

于海月,张淇,卢兵,等.碱熔分解-电感耦合等离子体质谱/发射光谱法测定黑土地地表基质中 34 种元素[J]. 岩矿测试, 2025, 44(3): 460-468. DOI: 10.15898/j.ykcs.202411290248.

YU Haiyue, ZHANG Qi, LU Bing, et al. Determination of 34 Elements in Ground Substrate of Black Soil by ICP-MS/OES with Alkali Melting Digestion[J]. Rock and Mineral Analysis, 2025, 44(3): 460–468. DOI: 10.15898/j.ykcs. 202411290248.

碱熔分解-电感耦合等离子体质谱/发射光谱法测定黑土地地表 基质中 34 种元素

于海月^{1,2},张淇^{1,2},卢兵^{1,2*},韩晨灏^{1,2},王桃源^{1,2},李明亮^{1,2} (1.中国地质调查局哈尔滨自然资源综合调查中心,黑龙江哈尔滨150086; 2.自然资源部哈尔滨黑土地地球关键带野外科学观测研究站,黑龙江哈尔滨150086)

摘要:黑土样品中有机质含量较高,通常在3%~10%之间,采用敞口酸溶、微波酸溶、高压密闭酸溶、 灰化法等传统的前处理方法并不能将样品完全消解,而不同的元素需要采用不同的消解和测定方法,导致 实验耗时长,操作过程复杂,分析成本过高,不适合多种元素同时测定。本文采用碳酸钠、四硼酸锂和偏 硼酸锂作为混合熔剂分解样品,并提出在酸提取剂中加入酒石酸的方法来提取样品,利用电感耦合等离子 体质谱法和发射光谱法 (ICP-MS/OES) 对黑土地地表基质中34种元素进行测定,解决了传统分析方法不适 合多元素同时测定和钽铪等元素易水解的问题。研究了不同提取体系对钽铪元素提取效果的影响,确定了 提取剂为30mL 盐酸-10mL 100g/L 酒石酸是最佳比例。方法检出限为0.01~50.13µg/g,测定结果能够满足 土壤分析要求。采用不同种类的土壤标准物质对本方法进行验证,各元素的测定值与标准值基本一致, 相对标准偏差 (RSD)为0.48%~4.53%,相对误差 (RE) 为-5.23%~4.85%,实际样品分析的相对标准偏差均 在0.16%~4.97%之间。

关键词: 电感耦合等离子体质谱法; 电感耦合等离子体发射光谱法; 酒石酸; 碱熔; 黑土

要点:

(1)为有效溶解有机质含量较高的样品,选用高熔点且非氧化性的碳酸钠、四硼酸锂和偏硼酸锂作为熔剂。(2)在酸提取剂中加入酒石酸与钽铪钨等易水解元素进行络合,获得了较理想的提取效果。

(3) 本文建立的方法与封闭酸溶法的测定结果基本一致,并且操作更加简单。

中图分类号: O657.63; S151.9 文献标识码: B

黑土地是世界上最肥沃的土壤,因其富含有机 质而呈黑色,故而得名^[1-3]。中国东北地区的黑土 地作为世界三大黑土区之一,是东北农业生产活动 的主要载体,其土壤环境质量严重影响着国家粮食产 率^[4-5]。但是,近年来由于全球自然环境变化和人 类的高强度开发利用,土壤酸化、耕作层变薄、有机 质含量下降等问题日益突出^[6-8]。地表基质是地球 表层孕育和支撑各类自然资源的基础物质,在 0~2m深度内具有环境指示意义的松散沉积物,包 括耕作层土壤、河流淤积层及风化残积物,用于分析 重金属迁移规律及生态风险^[9]。因此,针对黑土地 地表基质开展调查,对于黑土保护与可持续发展具 有极其重要的意义。

黑土地地表基质样品中有机质含量较高,通常

收稿日期: 2024-11-29; 修回日期: 2025-03-21; 接受日期: 2025-04-07; 网络出版日期: 2025-04-21 基金项目: 中国地质调查局地质调查项目"战略性矿产勘查岩矿测试技术支撑 (哈尔滨中心)"(DD20243204) 第一作者: 于海月,硕士研究生,助理工程师,主要从事地质实验测试工作。E-mail: yuhaiyue@mail.cgs.gov.cn。 通信作者: 卢兵,正高级工程师,主要从事地质实验测试工作。E-mail: LUBING_007@163.com。

在3%~10%之间^[10],常见的前处理方法如酸 溶^[11]不能将其完全消解,因此容易导致测试结果 不准确。目前,电感耦合等离子体质谱法 (ICP-MS) 和电感耦合等离子体发射光谱法 (ICP-OES) 是主要 的分析方法^[12-14],在地表基质调查分析中具有广阔 的应用前景,但样品前处理方法的建立是应用 ICP-OES 和 ICP-MS 进行分析的关键^[15-16]。目前常见 的样品前处理方法有敞口酸溶^[11]、微波酸溶^[17-19]、 高压密闭酸溶^[20-21]、灰化法^[22]等,其中,酸溶法 不能将有机质含量较高的黑土样品完全消解,不同 的元素需要采用不同的消解方法和测定方法。例如, 谢旭等^[23]用微波消解 ICP-MS 法测定黑土地地表 基质调查样品中的碘,杜贯新^[24]用 ICP-MS 法测定 黑土中的重金属元素,均对已有的消解体系进行了 优化改进,测试结果相对于传统方法更为准确,但其 分析测试指标局限性较大,碘和重金属的测试需要 采用不同的前处理方法,不适合多种类型元素的同 时测试,导致实验室测试指标较多时,实验耗时长, 分析成本高。灰化法虽然比较适合分解有机质含量 较高的样品,但是在灰化过程中容易导致钾、钠、铝 等元素损失,也不适用于多种元素的同时测定^[25]。 因此,研究一种简便、快速的黑土地地表基质成分分 析前处理方法是十分有必要的。

第3期

碱熔法近年来得到不断发展,在样品分析中发 挥了不可或缺的作用^[26-28]。碱熔法的熔样时间短, 样品分解比较完全,而且能够分解一些难溶于酸的 酸性氧化物及硅酸盐,但是由于强碱性熔剂(如过氧 化钠、氢氧化钠)引入盐分较高,易堵塞喷雾器,溶液 需高度稀释因而检出能力降低^[29-30],故在实际应用 中无法满足同时测定多种成分的需求。因此,本文 参考了国家标准方法《土壤和沉积物11种元素的 测定碱熔-电感耦合等离子体发射光谱法》 (HJ 974—2018)和前人研究^[31-32],选择碳酸钠、偏 硼酸锂、四硼酸锂作为混合熔剂,该混合熔剂是具有 高熔点的非氧化性熔剂,对样品有很强的分解能力, 克服了传统的酸溶方法带来的有机质消解不完全的 问题,并且采用浓盐酸与能抑制钽、铪元素水解的酒 石酸为提取剂,确定最佳提取条件,利用 ICP-MS 法 测定黑土地表基质样品中铋钴钽铪钨钼铊铀等稀有 金属以及稀土元素共 24 种元素,同时利用 ICP-OES 法对样品中锰钛钡锶钾钠钙镁铝铁 10 种金属元素 进行测定, 拟为黑土地地表基质调查提供一种快速、 准确的分析方法。

1 实验部分

1.1 仪器与工作条件

1.1.1 电感耦合等离子体质谱仪

X Series II 电感耦合等离子体质谱仪 (美国 ThermoFisher 公司)。工作条件为: 射频功率 1200W; 冷却气流速 15L/min; 辅助气流速 0.8L/min; 雾化气 流速 0.79L/min; 雾化器压力 2.5×10⁵Pa; 驻留时间 10ms。待测元素同位素分别为: ²⁰⁹Bi、⁵⁹Co、¹³⁹La、 ¹⁴⁰Ce、¹⁴¹Pr、¹⁴⁶Nd、¹⁴⁷Sm、¹⁵³Eu、¹⁵⁷Gd、¹⁵⁹Tb、¹⁶³Dy、 ¹⁶⁵Ho、¹⁶⁶Er、¹⁶⁹Tm、¹⁷²Yb、¹⁷⁵Lu、⁸⁹Y、¹⁸¹Ta、¹⁷⁸Hf、 ¹⁸²W、⁹⁸Mo、²³²Th、²⁰⁵Tl、²³⁸U。

1.1.2 电感耦合等离子体发射光谱仪

Plasma 3000 电感耦合等离子体发射光谱仪 (钢 研纳克检测技术股份有限公司)。工作条件为: 射频 功率 1250W; 冷却气流速 15L/min; 辅助气流速 0.2L/min; 载气流速 0.8L/min; 检测器温度-40℃; 光谱室温度 38℃。待测元素分别为: Mn、Ti、Ba、Sr、 K、Na、Ca、Mg、Al、Fe。

1.2 实验样品和标准物质

选取不同有机质含量 (3.57%、1.31%、2.6%) 的 土壤成分分析国家一级标准物质 GBW07978、 GBW07979、GBW07980(中国地质科学院地球物理 地球化学勘查研究所研制) 和采集自黑龙江省海伦 市的黑土地地表基质样品 (编号为 S1、S2、S3、S4、 S5) 作为实验样品,共同作为本方法准确度与精密度 的验证样品。采集后的样品经风干、研磨和过 200 目筛后使用。

1.3 标准溶液和主要试剂

多元素混合标准储备溶液:BWT30063-10-N-100(钇镧铈镨钕钐铕钆铽镝钬铒铥镱镥钍,各元素质 量浓度分别为 10mg/L); NCS148589(铋钴铪钽钨钼 铊,各元素质量浓度分别为 100μg/mL); BWJ4002-2016(铀,质量浓度为 1000μg/mL), NCS141835(锰钛 钡锶,各元素质量浓度分别为 10mg/L); NCS148701 (钾钠钙镁铝铁,各元素质量浓度分别为 100μg/mL)。

在样品消解过程中,引入偏硼酸锂会导致溶液 中盐基体影响较为严重,因此参考了国家标准方法 《土壤和沉积物 11 种元素的测定 碱熔-电感耦合等 离子体发射光谱法》(HJ 974—2018) 和前人研究成 果^[33],在配制空白溶液和标准溶液时,采取与样品 溶液相同的基体进行匹配,以降低基体效应。移取 适量的多元素混合标准储备液,用与样品溶液相同 的基体溶液逐级稀释成多元素混合标准溶液系列,

— 461 —

分别为混标 1、混标 2、混标 3。具体元素及浓度列 于表 1。

表1 标准溶液系列浓度

 Table 1
 Different concentrations of standard solution series.

混合标准储备 溶液编号	待测元素	浓度
	Bi, Co, La, Ce, Pr, Nd,	
调标 1	Sm, Eu, Gd, Tb, Dy, Ho,	0, 1, 2, 5, 10,
16275 1	Er, Tm, Yb, Lu, Y, Ta,	20, 50µg/L
	Hf, W, Mo, Th, Tl, U	
混标 2	Mp Ti Ba Sr	0, 0.1, 1, 2, 5,
1867/17 2	Ivili, 11, Da, Si	10, 20mg/L
混标 3	K Na Ca Mg Al Fe	0, 1, 10, 20, 50,
105705 3	K, Ma, Ca, Mg, Al, PC	100, 200mg/L

熔剂:碳酸钠(天津市光复科技发展有限公司)、 四硼酸锂(洛阳特耐实验设备有限公司)、偏硼酸锂 (上海麦克林生化科技股份有限公司)。

浸提剂:酒石酸(天津市科密欧化学试剂有限公司)、浓盐酸(天津市风船化学试剂科技有限公司)。

1.4 实验方法

1.4.1 样品前处理

在铂金坩埚底部加入少量碳酸钠垫底,称取 1.0g碳酸钠、0.1g四硼酸锂和 0.4g偏硼酸锂混匀制 成熔剂,再依次向坩埚中加入约 1g混合熔剂和 0.2g 样品 (精确至 0.0001g),最后放入剩余的熔剂,使其铺 匀在混合物的表面。盖上坩埚盖,将铂金坩埚置于 马弗炉中,升温至 1000℃,保持 30min,停止加热后 立即将铂金坩埚置于已盛有 100mL 热水的烧杯中, 熔融物出现裂纹并逐渐与坩埚脱离,洗出坩埚,在烧 杯中加入 10mL 100g/L 酒石酸和 30mL 浓盐酸,待熔 融物完全溶解,将其转移至 500mL 容量瓶中,定容, 待测。

1.4.2 电感耦合等离子体质谱分析

在使用 ICP-MS 进行元素检测的过程中,由于

土壤样品基体复杂多变,导致样品检测结果受基体效应影响较大,因此选择低质量数的¹⁰³Rh 和高质量数的¹⁶⁸Re 作为内标元素,以校正基体效应。

1.4.3 电感耦合等离子体发射光谱分析

针对 ICP-OES 存在的光谱干扰和非光谱干扰, 一般通过选择干扰少且信噪比高的分析谱线来降低 光谱干扰。通过扫描标准溶液与空白溶液,综合分 析各元素谱线的强度、稳定性及干扰,优化选择各元 素的分析谱线参数,测定波长为: Mn 257.610nm, Ti 334.940nm, Ba 455.403nm, Sr 460.733nm, K 766.490 nm, Na 818.326nm, Ca 315.887nm, Mg 285.213nm, Al 396.152nm, Fe 259.940nm。

2 结果与讨论

2.1 样品浸提条件的优化

为了获得更准确的分析结果,选择了可以使更 多元素以稳定可溶性形式存在的碱熔法,但钽、铪元 素较易水解^[34],如果仅使用浓盐酸提取,会致使测 试结果不准确。因此,为了解决这个问题,在酸提取 剂中加入一定量的酒石酸与易水解元素形成络合物 来抑制水解,达到较为理想的提取效果。

为了确定酒石酸的最佳用量,选择盐酸用量递 减、酒石酸用量递增的方式进行验证。分别选用了 4种不同比例的提取剂:40mL盐酸、30mL盐酸 -10mL 100g/L酒石酸溶液、20mL盐酸-20mL 100g/L酒石酸溶液、10mL盐酸-30mL 100g/L酒石 酸溶液,对熔融完成的三种地球化学调查样品标准 物质 GBW07978、GBW07979、GBW07980进行提取 测试。以较易水解的钽、铪元素为例,表2列出了在 使用不同量的提取剂提取样品的条件下,三种标准 物质中钽、铪元素的测定值。由数据分析可知,以 40mL盐酸作提取剂时,测定值较标准值整体偏低, 可能是由于此种情况下钽、铪水解较为严重;另外在 三种加入不同含量酒石酸的体系中,30mL盐酸

表 2 不同提取体系中土壤标准物质钽、铪元素的测定结果

Table 2 Analytical results of Ta and Hf in soil reference materials with different extraction systems.

标准物质 编号		Ta	含量 (µg/g)		Hf 含量 (µg/g)						
	标准估	40mL 盐酸	盐酸-酒石酸体积比			标准估	40mL 盐酸	盐酸-酒石酸体积比			
	你们出租	测定值	3:1	1:1	1:3	你们出租	测定值	3:1	1:1	1:3	
GBW07978	1.26	0.99	1.24	1.16	1.11	8.3	7.01	8.21	8.0	7.9	
GBW07979	1.27	1.03	1.26	1.19	1.15	7.3	6.35	7.23	6.96	6.82	
GBW07980	2.63	2.01	2.55	2.45	2.36	7.9	6.15	7.83	7.66	7.45	

-10mL 100g/L 酒石酸溶液体系的测试结果最接近标准值, 钽、铪元素的提取效果最好。因此, 本实验选择 30mL 盐酸-10mL 100g/L 酒石酸溶液作为提取剂。

2.2 标准曲线和方法检出限

分别对多元素混合标准溶液进行测定,以各元 素的质量浓度为横坐标,对应的信号值为纵坐标绘 制标准曲线,所得标准曲线的相关系数均不小于 0.9990。另外,对照样品实验方法连续平行测定空白 样品 11 次,以计算结果的 3 倍标准偏差为检出限, 34 种元素检出限为 0.01 ~ 50.13µg/g(表 3),能够满足 《地质矿产实验室测试质量管理规范》(DZ/T 0130—2006)的分析要求。

2.3 方法精密度和准确度

为了验证本方法的精密度和准确度,按照实验 方法对地球化学调查土壤标准物质 GBW07978、 GBW07979、GBW07980 平行测定 11 次,相对误差 (RE)均在-5.23%~4.85% 之间,说明该方法准确可 靠;各元素测定值的相对标准偏差 (RSD)均在 0.48%~4.53% 之间(表 4)。说明该方法精密度和重 现性较好,适用于黑土土壤样品的批量分析。

2.4 实际样品分析

为验证本方法在实际工作中的可行性,对黑龙 江海伦地区黑土地地表基质 3 件样品 (S1、S2、S3) 平行 测定 10 次,各成分测定值的 RSD 均在 0.16%~4.97%之间,小于 5%(表 5),表明该方法精密

表 3 线性回归方程、相关系数和方法检出限

Table 3	Linear regression equation	, correlation coefficient a	and detection limit of the method.

测空方法	一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一一	线性范围	线树同山古母	相关系数	方法检出限
侧走力伝	九系	(mg/L)	线住西归力性	(R^2)	$(\mu g/g)$
	Bi	0.1	<i>y</i> =40.351 <i>x</i> +0.239	0.9998	0.10
	Со	0.1	y=0.015x-0.0055	0.9996	0.03
	La	0.1	y=0.0471x-0.102	0.9999	0.10
	Ce	0.1	y=0.0505x-0.1016	0.9999	0.06
	Pr	0.1	y=0.0741x-0.1802	0.9999	0.01
	Nd	0.1	<i>y</i> =0.0272 <i>x</i> -0.0517	0.9999	0.02
	Sm	0.1	y=0.0259x-0.1029	0.9997	0.02
	Eu	0.1	<i>y</i> =0.0513 <i>x</i> -0.028	0.9998	0.01
	Gd	0.1	y=0.0236x+0.0011	0.9998	0.02
	Tb	0.1	<i>y</i> =0.1101 <i>x</i> -0.3521	0.9999	0.01
	Dy	0.1	y=0.0328x+0.0183	0.9997	0.02
ICP-MS	Но	0.1	y=0.1182x-0.4174	0.9999	0.01
101-1015	Er	0.1	<i>y</i> =0.0395 <i>x</i> +0.0808	0.9999	0.01
	Tm	0.1	y=0.121x-0.0324	0.9997	0.02
	Yb	0.1	<i>y</i> =0.0429 <i>x</i> +0.0003	0.9999	0.01
	Lu	0.1	y=0.1457x-0.5858	0.9999	0.01
	Y	0.1	<i>y</i> =0.0347 <i>x</i> +0.0356	0.9998	0.06
	Та	0.1	<i>y</i> =0.0347 <i>x</i> +0.0356	0.9998	0.01
	Hf	0.1	$y=0.0003x+1.26\times10^{-6}$	0.9998	0.03
	W	0.1	y=95.71x+1.0292	0.9998	0.21
	Mo	0.1	y=0.0227x-0.0021	0.9999	0.02
	Th	0.1	y=0.086x-0.0021	0.9993	0.02
	Tl	0.1	<i>y</i> =0.0451 <i>x</i> -0.0121	0.9997	0.01
	U	0.1	y=0.1328x-0.0523	0.9995	0.03
	Al_2O_3	100	<i>y</i> =7.484 <i>x</i> +198.657	0.9996	41.04
	TFe_2O_3	100	<i>y</i> =327.867 <i>x</i> +47.898	0.9998	37.59
	MgO	100	<i>y</i> =139249.452 <i>x</i> +31935.310	0.9993	4.74
	CaO	100	<i>y</i> =3914.427 <i>x</i> +2990.342	0.9992	17.76
ICP-OFS	Na ₂ O	100	<i>y</i> =327.867 <i>x</i> +47.898	0.9998	50.13
101-025	K_2O	100	<i>y</i> =1875.342 <i>x</i> +2452.084	0.9994	38.52
	Mn	10	<i>y</i> =34.057 <i>x</i> +459.306	0.9994	1.32
	Ti	10	<i>y</i> =37.370 <i>x</i> -644.631	0.9994	2.46
	Ba	10	<i>y</i> =541247.043 <i>x</i> +1325.417	0.9998	0.63
	Sr	10	y=11023.809x+670.300	0.9998	0.15

表 4 方法精密度和准确度测试结果

Table 4 Precision and accuracy tests of the method.

		GBW	07978			GBW	07979		GBW07980			
元素	标准值	测定值	相对误差	RSD	标准值	测定值	相对误差	RSD	标准值	测定值	相对误差	RSD
	(µg/g)	(µg/g)	(%)	(%)	(µg/g)	(µg/g)	(%)	(%)	(µg/g)	(µg/g)	(%)	(%)
Bi	0.38	0.37	-1.70	1.72	0.31	0.30	-4.81	2.91	301	301.66	0.22	1.14
Co	16	16.33	2.09	1.7	15.5	15.57	0.47	1.10	9.5	9.63	1.40	1.43
La	41	39.68	-3.22	0.86	37	37.48	1.29	0.90	55	57.52	4.58	1.18
Ce	82	80.74	-1.54	1.23	70	72.83	4.04	1.01	111	115.20	3.78	0.76
Pr	9.6	9.36	-2.47	1.08	8.5	8.44	-0.76	1.22	12.5	13.02	4.15	1.21
Nd	37	35.07	-5.23	1.53	33.2	34.04	2.54	0.56	45.9	47.05	2.51	0.94
Sm	7.1	7.12	0.22	2.14	6.3	6.27	-0.41	0.96	9.0	9.06	0.62	0.99
Eu	1.4	1.39	-0.86	1.56	1.4	1.41	0.60	1.37	1.27	1.23	-3.44	1.63
Gd	6.2	6.14	-0.92	1.32	5.8	5.99	3.35	1.72	8.0	8.06	1.07	1.17
Tb	1	1.03	2.51	2.03	0.94	0.91	-3.41	1.87	1.34	1.39	3.39	2.50
Dy	5.6	5.65	0.84	0.48	5.2	5.24	0.84	1.45	7.8	7.86	0.78	1.38
Но	1.12	1.10	-1.78	1.51	1.04	1.05	1.16	1.64	1.55	1.57	1.33	0.97
Er	3.2	3.06	-4.39	1.95	2.9	2.95	1.80	0.73	4.6	4.82	4.8	1.14
Tm	0.51	0.52	2.57	1.37	0.46	0.47	2.27	2.34	0.75	0.77	2.09	1.17
Yb	3.2	3.35	4.80	1.78	2.9	2.94	1.30	1.17	5.1	5.06	-0.69	1.14
Lu	0.5	0.51	2.51	0.96	0.43	0.44	2.83	1.72	0.78	0.78	0.14	1.94
Y	29.9	30.05	0.51	0.78	28.6	28.63	0.12	1.08	43.7	45.49	4.09	1.11
Та	1.26	1.27	0.57	1.88	1.27	1.26	-1.17	0.94	2.63	2.65	0.89	1.03
Hf	8.3	8.08	-2.65	1.20	7.3	7.23	-0.92	1.5	7.9	8.03	1.61	1.48
W	2.1	2.20	4.85	1.02	1.8	1.82	1.31	1.54	164	156.9	-4.33	1.38
Mo	0.68	0.65	-3.80	3.81	1.11	1.11	0.03	2.49	12.2	12.79	4.83	1.14
Th	13.6	13.53	-0.55	1.82	11.2	11.2	-1.05	1.30	23.9	24.58	2.83	0.99
T1	0.73	0.74	1.96	1.56	0.58	0.60	4.50	1.69	2.13	2.11	-1.02	1.13
U	3.1	3.05	-1.68	1.43	2.1	2.08	-1.11	1.17	15.6	15.36	-1.55	0.85
Al_2O_3	148400	143880	-3.05	1.27	132100	135064	2.24	1.71	127000	128578	1.24	0.79
TFe_2O_3	51300	53188	3.68	2.93	49800	49198	-1.21	4.36	82900	85404	3.02	2.18
MgO	12400	11998	-3.24	3.21	20700	21485	3.79	3.89	9200	9441	2.62	1.76
CaO	13900	14023	0.88	1.69	21800	22191	1.79	1.68	11800	12137	2.86	1.89
Na ₂ O	17300	17840	3.12	2.98	19500	20206	3.62	4.53	1300	1335	2.69	3.56
K_2O	27200	27411	0.78	2.00	25100	25720	2.47	2.84	18300	17964	-1.84	2.05
Mn	923	908	-1.60	2.35	674	660	-2.38	1.48	1675	1740	3.86	2.47
Ti	4760	4742	-0.37	2.13	4570	4616	1.00	1.98	4100	4138	0.93	1.17
Ba	639	641	0.38	3.17	679	674	-0.79	4.28	255	248	-2.71	3.49
Sr	186	182	-2.40	2.51	151	154	1.82	2.70	26.3	26.45	0.55	1.69

度较好。利用本方法和封闭酸溶法^[35]对两个实际 样品 (S4、S5)进行处理和测定,两种方法得到的稀 土元素以及易水解元素钽、铪的测试结果基本一致 (表 6)。以上结果均表明本方法能够满足实验室黑土 土壤日常检测要求。

3 结论

— 464 —

采用碳酸钠、偏硼酸锂、四硼酸锂作为熔剂,减 少了碱使用量并使样品完全消解。同时在浓盐酸提 取剂中引入适量酒石酸,以抑制钽、铪、钨等元素的 水解,建立了 ICP-MS 和 ICP-OES 相结合的分析方 法,实现了黑土地地表基质中 34 种元素的同时测定。 该方法可检测的元素种类较多,方法的精密度、准确 度、检出限均满足土壤样品分析要求,能够应用于实 际土壤样品的检测。

该方法前处理简单快速,熔样只需 30min,与常 规方法相比,缩短了样品处理时间,提高了工作效 率,有效地解决了有机质难消解和钽铪元素易水解 的问题,实现了黑土地地表基质中主量和微量元素 的同时测定,为土壤样品的批量分析提供了思路。 但本方法是基于黑土土壤样品,对其他类型样品 (如褐土、红壤等)批量分析还需进一步验证。

表 5 黑土地地表基质样品中 34 种元素分析结果 (n=10)

实际样品 S		品 S1	实际样	品 S2	实际样品 S3			实际样品 S1		实际样品 S2		实际样品 S3	
元素	测定值	RSD	测定值	RSD	测定值	RSD	元素	测定值	RSD	测定值	RSD	测定值	RSD
	$(\mu g/g)$	(%)	$(\mu g/g)$	(%)	(µg/g)	(%)		$(\mu g/g)$	(%)	$(\mu g/g)$	(%)	(µg/g)	(%)
Bi	2.89	0.16	0.24	3.74	0.17	2.36	Та	0.91	2.19	1.08	2.13	1.57	1.17
Co	11.16	1.19	13.64	1.52	11.45	1.46	Hf	9.13	1.46	6.87	2.09	6.73	1.66
La	21.57	1.37	26.71	2.17	25.24	1.86	W	0.87	3.62	1.28	2.00	1.08	1.57
Ce	40.74	1.55	50.85	1.57	47.67	1.82	Mo	0.38	4.97	0.52	3.61	0.78	2.77
Pr	4.66	2.17	5.75	2.35	5.57	2.54	Th	5.48	2.42	8.90	1.10	6.17	1.62
Nd	17.39	1.37	21.03	2.00	20.67	1.32	T1	0.40	2.49	0.55	1.92	0.46	2.70
Sm	3.05	2.55	3.75	1.93	3.86	2.28	U	1.36	2.00	1.58	2.00	2.52	1.63
Eu	0.70	2.40	0.83	1.63	0.86	2.20	Al ₂ O ₃	130291	1.57	131927	1.40	135236	1.15
Gd	3.17	2.17	3.75	2.85	3.88	2.38	TFe ₂ O ₃	52914	3.72	57903	2.55	47205	3.30
Tb	0.44	3.67	0.55	1.26	0.57	4.56	MgO	10160	3.26	11095	2.95	11263	2.11
Dy	2.38	2.26	3.20	2.18	3.17	2.25	CaO	9997	3.65	9990	2.61	11334	2.14
Но	0.48	2.25	0.65	1.83	0.62	2.26	Na ₂ O	17326	3.97	17783	2.7	20010	3.49
Er	1.31	1.99	1.83	2.03	1.72	2.46	K ₂ O	28325	3.2	27825	2.35	29015	1.91
Tm	0.20	3.15	0.29	2.22	0.27	1.92	Mn	1792	2.46	1134	2.29	2605	2.66
Yb	1.31	1.77	1.92	1.70	1.72	1.93	Ti	5331	1.58	5676	2.02	5905	1.84
Lu	0.21	1.35	0.30	2.01	0.26	2.28	Ba	863	2.52	686	2.53	968	2.95
Y	16.56	1.42	22.72	1.62	21.92	1.74	Sr	143	1.73	140	2.59	197	2.21

Table 5 Analytical results of 34 elements in ground substrate samples of black soil (n=10).

表 6 实际样品采用不同消解方法 34 种元素分析结果

Table 6 Analytical results of 34 elements in actual samples with different digestion methods.

	实际样,	品 S4	实际样	昻 S5		实际样。	品 S4	实际样品 S5	
元素	本方法 (碱熔法)	封闭酸溶法	本方法 (碱熔法)	封闭酸溶法	元素	本方法 (碱熔法)	封闭酸溶法	本方法 (碱熔法)	封闭酸溶法
	测定值 (μg/g)	测定值 (μg/g)	测定值 (μg/g)	测定值 (µg/g)		测定值 (μg/g)	测定值 (μg/g)	测定值 (μg/g)	测定值 (μg/g)
Та	0.542	0.577	8.3	7.01	Gd	2.005	2.089	6.004	5.818
Hf	5.483	5.964	7.9	6.15	Tb	0.351	0.345	0.958	0.881
La	13.19	13.59	45.48	44.81	Dy	1.846	1.974	5.697	5.351
Ce	23.88	25.39	86.46	84.29	Ho	0.403	0.424	1.108	1.051
Pr	3.169	3.358	9.75	9.554	Er	1.151	1.214	3.219	3.154
Nd	12.59	13.02	37.07	36.31	Tm	0.203	0.21	0.504	0.499
Sm	2.334	2.46	6.89	6.731	Yb	1.345	1.379	3.182	3.185
Eu	0.588	0.633	1.31	1.269	Lu	0.213	0.234	0.499	0.485

Determination of 34 Elements in Ground Substrate of Black Soil by ICP-MS/OES with Alkali Melting Digestion

YU Haiyue^{1,2}, ZHANG Qi^{1,2}, LU Bing^{1,2*}, HAN Chenhao^{1,2}, WANG Taoyuan^{1,2}, LI Mingliang^{1,2}

(1. Harbin Natural Resources Comprehensive Survey Center, China Geological Survey, Harbin 150086, China;

2. Harbin Observation and Research Station of Earth Critical Zone in Black Soil, Ministry of Natural Resources, Harbin 150086, China)

HIGHLIGHTS

- In order to effectively dissolve samples with high organic matter content, NaCO₃, LiBO₂ and Li₂B₄O₇ with high melting point and non-oxidizing properties were selected as fluxes.
- (2) Tartaric acid was added to the acid extraction agent and complexed with Ta, Hf, W and other easily hydrolyzable elements, and the ideal extraction effect was obtained.
- (3) The method established here is basically consistent with the determination results of the blocked acid solution method, and the operation is simpler.

ABSTRACT: The organic matter content in black soil samples is high, typically ranging from 3% to 10%. Traditional pretreatment methods such as open acid dissolution digestion, microwave acid dissolution and digestion, high-pressure closed acid dissolution and digestion, and ashing method cannot completely digest the sample, and different elements need to adopt different digestion and determination methods, resulting in long experiments, complex operation processes, high analysis costs, and are not suitable for the simultaneous determination of multiple elements. Sodium carbonate, lithium tetraborate and lithium metaborate were selected as mixed fluxes to digest the samples, and a method of adding tartaric acid to the acid extractor was proposed to extract the samples. 34 elements in the ground substrate of black soil were determined by inductively coupled plasma-mass spectrometry/optical emission spectrometry (ICP-MS/OES). The traditional black soil sample analysis method is not suitable for the simultaneous determination of multiple elements and the easy hydrolysis of elements such as tantalum and hafnium, while ICP-MS/OES is suitable. The effects of different extraction systems on the extraction of tantalum and hafnium were studied, and the optimal ratio of 30mL of hydrochloric acid and 10mL of 100g/L tartaric acid was determined. On this basis, the detection limit of the samples was $0.010-50.13\mu g/g$ by the experimental method of control samples, and the results satisfied the requirements of soil analysis and detection. At the same time, different kinds of soil reference materials were selected for practical testing, and the measured values of each element were basically consistent with the standard values, with the relative standard deviations (RSDs) between 0.48%-4.53% and the relative errors between -5.23%-4.85%. In addition, to further verify the feasibility of the method, the relative standard deviations were all between 0.16%-4.97% when applied to actual samples.

KEY WORDS: inductively coupled plasma-mass spectrometry; inductively coupled plasma-optical emission spectrometry; tartaric acid; alkali fusion; black soil

参考文献

[1] 刘航,李瑞红,李胜荣,等.松辽平原典型黑土区不同 土壤结构的古气候背景与黑土成因机制[J].地质 通报,2024,43(5):766-778.

> Liu H, Li R H, Li S R, et al. Paleoclimate background and genetic mechanisms of different soil structures in typical black soil areas of Songliao Plain[J]. Geological Bulletin of China, 2024, 43(5): 766–778.

[2] 刘顺国,徐英德,裴久渤,等.以土壤普查成果助推黑
 土地科学保护与利用[J].土壤通报,2024,55(4):
 1185-1190.

Liu S G, Xu Y D, Pei J B, et al. Promoting the scientific protection and utilization of black soil through the third national soil survey results[J]. Chinese Journal of Soil Science, 2024, 55(4): 1185–1190.

[3] 高佳,朱耀辉,赵荣荣.中国黑土地保护:政策演变、现 实障碍与优化路径[J].东北大学学报(社会科学版), 2024,26(1):82-89.

> Gao J, Zhu Y H, Zhao R R. Black soil protection in China: Policy evolution, realistic obstacles and optimization paths[J]. Journal of Northeastern University (Social Science), 2024, 26(1): 82–89.

[4] 马婧萱,崔晓丹,梁强,等.东北典型黑土富硒区土壤
 与作物硒含量特征及影响因素——以海伦市为例[J].
 土壤, 2024, 56(5): 1101-1110.

Ma J X, Cui X D, Liang Q, et al. Characteristics and influencing factors of Se content of soil and crop in Serich areas of isohumosols (mollisols) in Northeast China: A case study of Hailun [J]. Soils, 2024, 56(5): 1101–1110.

- [5] 马瑞,田芷源,赵艳,等.东北黑土典型水蚀区农田侵 蚀与退化特征[J].水土保持学报,2024,38(5):71-81.
 Ma R, Tian Z Y, Zhao Y, et al. Characteristics of erosion and degradation for farms and sin the typical water erosion area of the black soil region, Northeast China[J]. Journal of Soil and Water Conservation, 2024, 38(5): 71-81.
- [6] 司佳雪, 董鑫, 张晓晴, 等. 基于 CiteSpace 的国内外土 填污染治理研究进展和发展动态分析[J]. 现代化工, 2024, 44(9): 17-23.
 Si J X, Dong X, Zhang X Q, et al. CiteSpace based

analysis on research progress and development trend of global soil contamination control[J]. Modern Chemical Industry, 2024, 44(9): 17–23.

- [7] Chai L, Zhou Y Q, Dong H K, et al. Soil contamination and carrying capacity across the Tibetan Plateau using structural equation models[J]. Environmental Pollution, 2023, 337: 122640.
- [8] 王萌, 俞磊, 秦璐瑶, 等. 土壤环境基准的科学问题与研究方法: 以 Cd 为例[J]. 地学前缘, 2024, 31(2): 147-156.

Wang M, Yu L, Qin L Y, et al. Scientific issues and research methods of soil environmental standards: A case study on cadmium[J]. Earth Science Frontiers, 2024, 31(2): 147–156.

[9] 陈龙, 郭海全, 尚晓雨, 等. 新乐市地表基质有机碳储 量变化趋势及影响因素[J]. 黑龙江农业科学, 2023(3): 34-40.

Chen L, Guo H Q, Shang X Y, et al. Change trend and determinants of organic carbon storage in ground substrate of Xinle City[J]. Heilongjiang Agricultural Sciences, 2023(3): 34–40.

- [10] 贾建丽, 于妍. 环境土壤学[M]. 北京: 化学工业出版社, 2022: 258.
 Jia J L, Yu Y. Environmental soil science[M]. Beijing: Chemical Industry Press, 2022: 258.
- [11] 唐碧玉,张征莲,谷娟平,等.四酸消解-电感耦合等离子体质谱法测定铝土矿中 18 种元素的含量[J].理化检验(化学分册),2024,60(9):857-864.
 Tang B Y, Zhang Z L, Gu J P. Determination of 18 elements in bauxite by inductively coupled plasma mass spectrometry with four-acid digestion[J]. Physical Testing and Chemical Analysis (Part B: Chemical Analysis), 2024, 60(9): 857-864.
- [12] Provete S C, Dalfior M B, Mantovaneli R, et al. Comparison of the performance of ICP-MS, CV-ICP-OES, and TDA AAS in determining mercury in marine sediment samples[J]. ACS Omega, 2024, 9(50): 49229–49238.
- [13] Baghaliannejad R, Aghahoseini M, Amini K M. Determination of rare earth elements in uranium materials by ICP-MS and ICP-OES after matrix separation by solvent extraction with TEHP[J]. Talanta, 2021, 222: 121509.
- [14] 孙梦荷,苏春风,方迪,等.电感耦合等离子体质谱 (ICP-MS)法在稀土分析中的应用进展[J].中国无机 分析化学, 2023, 13(9): 939-949.
 Sun M H, Su C F, Fang D, et al. Application progress of inductively coupled plasma mass spectrometry (ICP-MS) in rare earth analysis[J]. Chinese Journal of Inorganic Analytical Chemistry, 2023, 13(9): 939-949.
- [15] Rito B, Almeida D, Coimbra C. Post-measurement compressed calibration for ICP-MS-based metal quantification in mine residues bioleaching[J]. Scientific Reports, 2022, 12(1): 16007.
- [16] Morrison C, Sun H, Yao Y, et al. Methods for the ICP-OES analysis of semiconductor materials[J]. Chemistry of Materials, 2020, 32(5): 1760–1768.
- [17] 王树加,郑巧清,卢思,等. 微波消解-电感耦合等离子体质谱 (ICP-MS) 法测定土壤 7 种金属元素 [J]. 中国

无机分析化学, 2023, 13(12): 1336-1341.

Wang S J, Zheng Q Q, Lu S, et al. Determination of seven metal elements in soil samples by inductively coupled plasma mass spectrometry with microwave digestion[J]. Chinese Journal of Inorganic Analytical Chemistry, 2023, 13(12): 1336–1341.

- [18] 官迪, 陈山, 田发祥, 等. 微波消解-电感耦合等离子体 质谱法测定有机肥中镉、砷、汞、铅、铬含量[J]. 中国 土壤与肥料, 2024(1): 226-231.
 Guan D, Chen S, Tian F X, et al. Determination of cadmium, arsenic, mercury, lead and chromium in organic fertilizer by microwave digestion combined with inductively coupled plasma mass spectrometry[J]. Soil and Fertilizer Sciences in China, 2024(1): 226-231.
- [19] 吴佳伦,罗霜,李思思,等.微波消解/石墨消解-ICP-MS测定土壤中的多种重金属[J].中国测试,2021,47(5):58-63.
 Wu G L, Luo S, Li S S, et al. Microwave direction/graphite direction ICP MS_determination_of

digestion/graphite digestion-ICP-MS determination of multiple heavy metals in soil[J]. China Measurement & Test, 2021, 47(5): 58–63.

[20] 王蕾, 于汀汀, 孙红宾, 等. 高压密闭酸溶-电感耦合等 离子体发射光谱法测定硼矿石中的硼[J]. 岩矿测试, 2024, 43(3): 468-475.
Wang L, Yu T T, Sun H B, et al. Boron analysis in boron ores by inductively coupled plasma-optical emission spectrometry with sealed acid digestion at high

pressure[J]. Rock and Mineral Analysis, 2024, 43(3): 468–475.

- [21] 邵坤, 余滔, 龚大兴. 高压密闭消解-ICP-AES 法测定沉积型稀土矿中 La、Ce、Pr、Nd、Y、Nb、Zr[J]. 稀土, 2024, 45(1): 87-94.
 Shao K, Yu T, Gong D X. Determination of La, Ce, Pr, Nd, Y, Nb and Zr in sedimentary rare-earth ore by ICP-AES with sealed digestion[J]. Chinese Rare Earths, 2024, 45(1): 87-94.
- [22] 郝志红,刘彬,杜雪苗,等.干法灰化-发射光谱法测定 植物样品中 B和 Sn[J].中国无机分析化学, 2023, 13(12):1437-1443.
 Hao Z H, Liu B, Du X M, et al. Determination of B, Sn

in plant samples by emission spectrometry with dry ashing[J]. Chinese Journal of Inorganic Analytical Chemistry, 2023, 13(12): 1437–1443.

[23] 谢旭, 卢兵, 赵文志, 等. 微波消解-电感耦合等离子体质谱 (ICP-MS) 法测定黑土地地表基质调查样品中碘[J]. 中国无机分析化学, 2024, 14(7): 980-985.
Xie X, Lu B, Zhao W Z, et al. Determination of iodine in black soil surface matrix survey samples by inductively

coupled plasma mass spectrometry (ICP-MS) with

第3期

microwave digestion[J]. Chinese Journal of Inorganic Analytical Chemistry, 2024, 14(7): 980–985.

- [24] 杜贯新. 松嫩黑土区西北部多流域黑土地球化学及重金属累积特征[D]. 大庆: 东北石油大学, 2023.
 Du G X. The geochemical and heavy metal accumulation characteristics of black soil in the northwest of Songnen black soil region were studied[D]. Daqing: Northeast Petroleum University, 2023.
- [25] 马树侠,高明飞. 微波消解/湿法灰化-电感耦合等离子体原子发射光谱法测定原油重油中 6 种金属元素的含量[J]. 理化检验(化学分册), 2022, 58(4): 481-484.
 Ma S X, Gao M F. Microwave digestion/wet ashing-inductively coupled plasma atomic emission spectrometry was used to determine the contents of six metal elements in crude oil and heavy oil[J]. Physical Testing and Chemical Analysis (Part B: Chemical Analysis), 2022, 58(4): 481-484.
- [26] 孙孟华,朱永晓,赵烨,等. 过氧化钠碱熔-电感耦合等 离子体质谱法测定地质样品中锆铌铪钽和稀土元素
 [J]. 冶金分析, 2024, 44(9): 35-44.
 Sun M H, Zhu Y X, Zhao Y, et al. Determination of zirconium, niobium, hafnium, tantalum and rare earth elements in geological sample by inductively coupled plasma mass spectrometry with sodium peroxide alkali fusion[J]. Metallurgical Analysis, 2024, 44(9): 35-44.
- [27] 席秀丽, 安婷婷. 电感耦合等离子体发射光谱法测定 高岭土中 8 项组分含量[J]. 分析仪器, 2024(3): 25-30.
 Xi X L, An T T. Determination of 8 components in kaolin by inductively coupled plasma emission spectrometry[J]. Analytical Instrumentation, 2024(3): 25-30.
- [28] 乔柱, 范广宇, 赵秀荣, 等. 碱熔酸浸-原子荧光光谱法 检测铬铁矿中砷和铅[J]. 分析试验室, 2023, 42(5): 613-617.
 Qiao Z, Fan G Y, Zhao X R, et al. Detection of arsenic

and lead in chromite by alkali fusion-acid dissolving atomic fluorescence spectrometry [J]. Chinese Journal of Analysis Laboratory, 2023, 42(5): 613–617.

- [29] 范俊楠,张钰,李国坡,等.基于不同前处理方式 ICP-MS 法测定土壤中 22 种金属元素[J]. 分析科学学报, 2024, 40(4): 454-460.
 Fan J N, Zhang Y, Li G P, et al. Determination of 22 metals in soil by ICP-MS based on different pretreatment methods[J]. Journal of Analytical Science, 2024, 40(4): 454-460.
- [30] 薛丁帅,李文君,王静,等. 地质样品中钻、镍的分析进展[J]. 岩石学报, 2023, 39(4): 1217-1232.
 Xue D S, Li W J, Wang J, et al. Advances in analysis for

cobalt and nickel in geological materials[J]. Acta Petrologica Sinica, 2023, 39(4): 1217–1232.

- [31] 郭琳, 于汀汀, 孙红宾, 等. 偏硼酸锂熔融-电感耦合等 离子体发射光谱法测定铍矿石中的铍及主量元素[J]. 岩矿测试, 2024, 43(2): 356-365.
 Guo L, Yu T T, Sun H B, et al. Determination of beryllium and major elements in beryllium ores by inductively coupled plasma-optical emission spectrometry with lithium metaborate fusion[J]. Rock and Mineral Analysis, 2024, 43(2): 356-365.
- [32] 袁润蕾,李小辉,于亚辉,等. 偏硼酸锂熔融-电感耦合等离子体发射光谱法测定有色金属矿中二氧化硅[J].现代化工, 2024, 44(S2): 391-396.
 Yuan R L, Li X H, Yu Y H, et al. Molten lithium metaborate-inductively coupled plasma emission spectrometry for determination of silicon dioxide in non-ferrous metal[J]. Modern Chemical Industry, 2024, 44(S2): 391-396.
- [33] 甘黎明,薛玉容,黄云杰,等.碱熔-基体匹配-电感耦合等离子体发射光谱 (ICP-OES) 法测定硬岩型锂矿中锂和铷[J/OL].中国无机分析化学 (2025-02-08)[2025-04-10]. http://kns.cnki.net/kcms/detail/11.6005.O6.20250208.1534.009.html.

Gan L M, Xue Y R, Huang Y J, et al. Determination of lithium and rubidium in lithium ore by inductively coupled plasma optical emission spectrometry (ICP-OES) with alkali melt and matrix matching[J/OL]. Chinese Journal of Inorganic Analytical Chemistry (2025-02-08)[2025-04-10]. http://kns.cnki.net/kcms/detail/11.60 05.06.20250208.1534.009.html.

- [34] 夏传波, 成学海, 姜云, 等. 密闭酸溶-电感耦合等离子体发射光谱/质谱法测定花岗伟晶岩中 32 种微量元素
 [J]. 岩矿测试, 2024, 43(2): 247-258.
 Xia C B, Cheng X H, Jiang Y, et al. Determination of 32 trace elements in granite pegmatite by inductively coupled plasma-optical emission spectrometry and mass spectrometry with closed acid dissolution[J]. Rock and Mineral Analysis, 2024, 43(2): 247-258.
- [35] 程祎,李志伟,于亚辉,等.高压密闭消解-电感耦合等 离子体质谱法测定地质样品中铌、钽、锆、铪和16种 稀土元素[J].理化检验(化学分册),2020,56(7): 782-787.

Cheng Y, Li Z W, Yu Y H, et al. Determination of Nb, Ta, Zr, Hf and 16 rare elements in granite pegmatite by inductively coupled plasma mass spectrometry with closed acid dissolution[J]. Physical Testing and Chemical Analysis (Part B: Chemical Analysis), 2020, 56(7): 782–787.